NIRT Barley SRS Bias Log

Tecator Model: _____ Serial # : _____ Location: _____

NOTE: Only results less than two weeks old and within a 5 ° F temperature range may be averaged. Results obtained after a bias adjustment, SRS replacement, or NIRT repair or replacement cannot be average with results obtained prior to the change (also, the 5-run correction can only be applied if five consecutive, valid runs are all positive or all negative). Every effort should be made to maintain the bias within the 0.06% Protein tolerance.

Date mm/dd/yy	Temp	**Protein Class:**						Oper Init	**Comments**
			Today's Bias	Multi-Run Averages			B*		
				2-Run	3-Run	5-Run			
		RH	±0.12	±0.09	±0.06	±0.04	R*		