

Grain Inspection Advisory Committee
April 7-8, 2015
Resolutions

1. The Advisory Committee commends FGIS for its work in implementing and testing of UGMA moisture meters; and recommends that for the Sample Information Monitoring System (SIMS) that FGIS provide on their website a listing by grains for the approved UGMA moisture meters the following information: the moisture standard deviation, \pm warning limit, and \pm action limit and the moisture range for which these limits are applicable.
2. The Advisory Committee recommends FGIS move forward with expanding the mycotoxin concentration ranges to the following:

Aflatoxin to 5 – 300 ppb
DON to 0.5 to 30 ppm
Fumonisin to 0.5 – 30 ppm

At the manufacturer's request FGIS will verify the supplemental analysis performance at the maximum range as established by the manufacturer.

3. The Advisory Committee recommends that FGIS develop reports based on Sample Information Monitoring System (SIMS) and Subjective Testing and Evaluation Process (STEP) data that reflects quality performance on a system-wide and official service provider level.